

GASTEC Instructions for No.132M Trichloroethylene Detector Tube

FOR SAFE OPERATION :

Read this manual and the instruction manual of your Gastec Gas Sampling Pump carefully.

⚠ WARNING :

1. Use only Gastec detector tubes in a Gastec pump.
2. Do not interchange or use non-Gastec parts or components in Gastec's detector tube and pump system.
3. The use of non-Gastec parts or components in Gastec's detector tube and pump system or use of a non-Gastec detector tube with a Gastec pump or use of a Gastec detector tube with a non-Gastec pump may result in property damage, serious bodily injury, and death; voids all warranties; and voids all performance and data accuracy guaranties.

⚠ CAUTION : If not observed, injuries to the operator or damage to the product may result.

1. When breaking the tube ends, keep away from eyes.
2. Do not touch the broken glass tubes, pieces and reagent with bare hand(s).
3. The sampling time represents the time necessary to draw the air sample through the tube.
The tube must be positioned in the desired sampling area for the entire sampling time or until the flow finish indicator indicates the end of the sample.

⚠ NOTES : For maintaining performance and reliability of the test results.


1. Use Gastec Gas Sampling Pump together with Gastec Detector Tubes only for the purposes specified in the instruction manual of the detector tube.
2. Use this tube within the temperature range of 0 - 40°C (32 - 104°F).
3. Use this tube within the relative humidity range of 0 - 90%.
4. This tube may be interfered with by the coexisting gases. Please refer to the "INTERFERENCES".
5. Shelf life and storage condition of the tube are marked on the label of the box of tube.

APPLICATION OF THE TUBE :

Use this tube for the detection of Trichloroethylene in air for the industrial areas and environmental atmospheric condition.

SPECIFICATION :

(As a result of Gastec's commitment to continued improvement, specifications are subject to change without notice.)



Measuring Range	2 - 5 ppm	5 - 100 ppm	100 - 250 ppm
Number of Pump Strokes	2	1	1/2
Correction Factor	0.4	1	2.5
Sampling Time	1 minute per pump stroke		30 seconds
Detecting Limit	0.4 ppm (n = 2)		
Colour Change	Yellow → Reddish purple		
Reaction Principle	Trichloroethylene is decomposed by nascent oxygen by oxidizing agent to liberate hydrogen chloride which discolours indicator to reddish purple.		

Coefficient of Variation : 10% (for 5 to 20 ppm), 5% (for 20 to 100ppm)

**** Shelf Life : Please refer to the Validity Date printed on the box of tube.**

**** Store the tubes in the refrigerator to keep at 10°C (50°F) or below.**

CORRECTION FOR TEMPERATURE, HUMIDITY AND PRESSURE :

Calibration of the Gastec detector Tube No.132M is based on a tube temperature of 20°C (68°F) and not the temperature of the gas being sampled, approximately 50% relative humidity and normal

atmospheric pressure.

Temperature : Since the tube is affected by the temperature , multiply the correction factor to the tube reading.

Tube Reading (ppm)	True Concentration (ppm)								
	0°C (32°F)	5°C (41°F)	10°C (50°F)	15°C (59°F)	20°C (68°F)	25°C (77°F)	30°C (86°F)	35°C (95°F)	40°C (104°F)
100	—	—	140	120	100	88	75	68	60
80	180	140	110	90	80	70	60	55	50
60	120	100	80	70	60	55	50	45	40
40	70	60	50	45	40	35	30	27.5	25
20	30	27.5	25	22.5	20	17.5	15	14	13
10	16	14	12	11	10	9	8	7	6
5	8	7	6	5.5	5	4.5	4	3.5	3

Humidity : No correction is required

Pressure : To correct for pressure, multiply by the tube reading by

$$\frac{\text{Tube Reading (ppm)} \times 1013 \text{ (hPa)}}{\text{Atmospheric Pressure (hPa)}}$$

MEASUREMENT PROCEDURE :

1. For leak tight check of the pump insert a fresh sealed detector tube into pump. Follow instructions provided with the pump operating manual.
2. Break tips off a fresh detector tube in the tube tip breaker of the pump.
3. Insert the tube securely into pump inlet with arrow (➔) on the tube pointing toward pump.
4. Make certain pump handle is all the way in. Align guide marks on pump body and handle.
5. Pull handle all the way out until it locks on 1 pump stroke (100ml). Wait 1 minute and confirm the completion of the sampling.
6. For lower than 5 ppm measurement, repeat the above sampling procedure one more time until the stain attains to the first calibration mark. For higher than 100 ppm measurement, prepare fresh tube and take 1/2 full pump strokes.
7. Read concentration at the interface of the stained-to unstained reagent.
8. If correction is needed, multiply the correction factors of temperature, pump strokes and pressure.

INTERFERENCES :

Substance	Concentration	Interference	Changes colour by itself to
Nitric Oxide, Nitrogen dioxide		No effect	No discoloration
Hydrogen chloride, Chlorine, Bromine		Plus error	Reddish purple
Acetone	≤ 200 ppm	No effect	No discoloration
Unsaturated Halogenated HCs		Plus error	Reddish purple
Aromatic hydrocarbons	≥ 100 ppm	Minus error	No discoloration

The table of this interference gases primarily expresses the interference of each coexisting gas in the gas concentration range , equivalent to the gas concentration. Therefore, the test result may be given positive result by the other substances not listed in the table. For more information is needed, please contact us or our distributors in your territory.

DANGEROUS AND HAZARDOUS PROPERTIES :

Threshold Limit Value-Time Weighted Average by ACGIH (2005) : 50 ppm (7 - 8 hours)

Threshold limit Value-Short Term Exposure Limit by ACGIH (2005) : 100 ppm (15 min)

DISPOSAL INSTRUCTION :

Reagent of the tube does not use toxic substances. When disposing the tube regardless of used or unused, follow the rules and regulations of the local government.

WARRANTY :

If you have any questions regarding gas detection and quality of the tubes, please feel free to contact your Gastec representatives.

Manufacturer : Gastec Corporation
8-8-6 Fukayanaka, Ayase-City, 252-1195, Japan

<http://www.gastec.co.jp/>

Telephone +81-467-79-3910 Facsimile +81-467-79-3979

IM00132ME1
Printed in Japan
06B1Z