# GASTEC Instructions for Carbon Monoxide Detector No. 1LC Tube

### FOR SAFE OPERATION:

Carefully read this manual and the instruction manual of your Gastec Gas Sampling Pump.

### / WARNING:

- 1. Use only Gastec detector tubes in a Gastec pump.
- 2. Do not interchange or use non-Gastec parts or components in Gastec's detector tube and pump system.
- 3. Using non-Gastec parts or components in Gastec's detector tube and pump system or using a non-Gastec detector tube with a Gastec pump or using a Gastec detector tube with a non-Gastec pump may damage your detector tube and pump system, or may cause serious injuries, or death to the end-user. It will also void all warranties, and guarantees regarding performance and data accuracy.

## CAUTION: If you do not observe the following precautions, you may suffer injuries or damage the product.

- 1. When breaking the tube ends, keep away from eyes.
- 2. Do not touch the broken glass tubes, broken pieces and reagent with bare hand(s).

### NOTES: For maintaining performance and reliability of the test results, observe the following.

- Use Gastec Gas Sampling Pump together with Gastec Detector Tubes only for the purposes specified in the instruction manual of the detector tube.
- 2. Use this tube within the temperature range of 0 40°C (32 104°F).
- 3. Use this tube within the relative humidity range of 0 90%.
- 4. This tube may be interfered with by the coexisting gases. Please refer to the table "INTERFERENCES" below.
- 5. The shelf life and storage condition of the tube are marked on the label of the tube box.

**APPLICATION OF THE TUBE:** Use this tube for detecting carbon monoxide in the air or in industrial areas and for determining the environmental atmospheric condition.

**SPECIFICATION:** (Because of Gastec's commitment to continued improvement, specifications are subject to change without notice.)



Measuring Range	1 - 30 ppm			
Number of Pump Strokes	1			
Stroke Correction Factor	1			
Sampling Time	4 minutes per pump stroke			
Detecting Limit	0.5 ppm (n = 1)			
Colour Change	White → Pale Pink			
Reaction Principle	CO + Palladium Salt (White) → Reaction Product			

Coefficient of Variation: 10% (for 1 to 10 ppm), 5 % (for 10 to 30 ppm)

- \*\* Shelf Life: Please refer to the validity date printed on the tube box.
- \*\* Store the tubes in a dark and cool place.

### CORRECTION FOR TEMPERATURE. HUMIDITY AND PRESSURE:

**Temperature:** Correct for temperature with the table below.

Temperature °C (°F)	0 (32)	5 (41)	10 (50)	15 - 40 (59 - 104)
Correction Factor	1.8	1.3	1.1	1.0

**Humidity:** No correction is required.

Pressure: To correct for pressure, use the formula below:

\_Tube Reading (ppm) ×1013 (hPa)

Atmospheric Pressure (hPa)

### MEASUREMENT PROCEDURE:

- For checking the leakage of the pump, insert a freshly sealed detector tube into the pump. Follow the instructions provided with the pump operation manual.
- 2. Break the tips off the fresh detector tube with the tube tip breaker in the pump.
- 3. Insert the tube into the pump inlet with the arrow ( ) on the tube pointing toward the pump.
- Make certain the pump handle is all the way in. Align the guide marks on the pump body with the guide marks on the handle
- 5. Pull the handle all the way out until it locks at one pump stroke (100 mL). Wait four minutes and confirm the completion of the sampling.
- 6. Read the concentration at the interface of the stained-to-unstained reagent immediately after the sampling.
- If necessary, multiply the readings by the correction factors of the temperature and atmospheric pressure respectively.

### INTERFERENCES:

Substance	Concentration	Interference	Interference gas only
Ammonia		No	No discolouration
Sulphur Dioxide	≥ 1/20	-	No discolouration
NItrogen Dioxide	≥ 1/20	+	No discolouration

This table of interference gases primarily expresses the interference of each coexisting gas in the concentration range, that is equivalent to the gas concentration. Therefore, the test result may show positive results due to other substances not listed in the table. If more information is needed, please contact us or our distributors in your territory.

### DANGEROUS AND HAZARDOUS PROPERTIES:

Threshold Limit Value-Time Weighted Average by ACGIH (2009): 25 ppm

Explosive range: 12.5 - 74%

**INSTRUCTIONS ON DISPOSAL:** The reagent of the tube does not use toxic substances. When disposing the tube regardless of whether it has been used or not, follow the rules and regulations of your local government.

**WARRANTY:** If you have any questions regarding gas detection and the quality of the tubes, please feel free to contact your Gastec representatives.

Manufacturer: Gastec Corporation 8-8-6 Fukayanaka, Ayase-City, 252-1195, Japan http://www.gastec.co.jp/

Telephone +81-467-79-3910 Fax +81-467-79-3979

IM001LCE3 Printed in Japan