

GASTEC

No.76

Instructions for Tetrahydrothiophene Detector Tube

FOR SAFE OPERATION :

Carefully read this manual and the instruction manual of your Gastec Gas Sampling Pump.

⚠ WARNING :

1. Use only Gastec detector tubes in a Gastec pump.
2. Do not interchange or use non-Gastec parts or components in Gastec's detector tube and pump system.
3. Using non-Gastec parts or components in Gastec's detector tube and pump system or using a non-Gastec detector tube with a Gastec pump or using a Gastec detector tube with a non-Gastec pump may damage your detector tube and pump system, or may cause serious injuries, or death to the end-user. It will also void all warranties, and guarantees regarding performance and data accuracy.

⚠ CAUTION : If you do not observe the following precautions, you may suffer injuries or damage the product.

1. When breaking the tube ends, keep away from eyes.
2. Do not touch the broken glass tubes, broken pieces and reagent with bare hand(s).
3. The sampling time represents the time necessary to draw the air sample through the tube. The tube must be positioned in the desired sampling area for the entire sampling time or until the flow finish indicator indicates the end of the sampling.


△ NOTES : For maintaining performance and reliability of the test results, observe the following.

1. Use Gastec Gas Sampling Pump together with Gastec Detector Tubes only for the purposes specified in the instruction manual of the detector tube.
2. Use this tube within the temperature range of 0 - 40°C (32 - 104°F).
3. Use this tube within the relative humidity range of 0 - 90%.
4. This tube may be interfered with by the coexisting gases. Please refer to the table "INTERFERENCES" below.
5. Do not expose the tube to direct sunlight. If the tube is exposed to direct sunlight for more than one hour, the reagent stains white for whole layer and lose its performance.
6. The shelf life and storage condition of the tube are marked on the label of the tube box.

APPLICATION OF THE TUBE :

Use this tube for detecting Tetrahydrothiophene in the air or in industrial areas and for determining the environmental atmospheric condition.

SPECIFICATION : (Because of Gastec's commitment to continued improvement, specifications are subject to change without notice.)



Measuring Range	1 - 10 ppm
Number of Pump Stroke	4
Stroke Correction Factor	1
Sampling Time	1 minute per pump stroke
Detecting Limit	0.25 ppm (n = 4)
Colour Change	Pink → Pale Yellow
Reaction Principle	C ₄ H ₈ S + MnO ₄ → Yellow Reaction Product

Coefficient of Variation : 10% (for 1 to 3 ppm), 5% (for 3 to 10 ppm)

**** Shelf Life :** Please refer to the validity date printed on the tube box.

**** Store the tubes in a dark and cool place**

CORRECTION FOR TEMPERATURE, HUMIDITY AND PRESSURE :

Temperature : Correct for temperature with the table below.

Temperature°C (°F)	0 (32)	5 (41)	10 (50)	15 (59)	20 (68)	25 (77)	30 (86)	35 (95)	40 (104)
Correction Factor	1.3	1.2	1.1	1.05	1.0	0.95	0.9	0.88	0.85

Humidity : No correction is required.

Pressure : To correct for pressure, use the formula below :

$$\frac{\text{Tube Reading (ppm)} \times 1013 \text{ (hPa)}}{\text{Atmospheric Pressure (hPa)}}$$

MEASUREMENT PROCEDURE :

1. For checking the leakage of the pump, insert a freshly sealed detector tube into the pump. Follow the instructions provided with the pump operation manual.
2. Break the tips off the fresh primary tube and analyser tube by snapping off each tube end in the tube tip breaker of the pump.
3. Connect the © marked ends with rubber tubing after snapping off each end.
4. Insert the analyser tube securely into pump inlet with the arrow ➔ on the tube pointing toward the pump.
5. Make certain the pump handle is all the way in. Align the guide marks on the pump body with the guide marks on the handle.
6. Pull the handle all the way out until it locks at one pump stroke (100 mL). Wait one minute and confirm the completion of the sampling. Repeat the above sampling procedure three more times.
7. Read the concentration level at the interface where the stained reagent meets the unstained reagent.
8. If necessary, multiply the readings by the correction factors of temperature and atmospheric pressure.

INTERFERENCES :

Substance	Concentration	Interference	Interference gas only
Hydrogen Sulphide	≤ 30 ppm	No	Pale pink for whole layer
Ethyl Mercaptan	≤ 1 ppm	No	Pale pink for whole layer
Ethylene	≤ 1000 ppm	No	Pale pink for whole layer
Propylene	≤ 500 ppm	No	Pale pink for whole layer
Methanol	≤ 300 ppm	No	Pale pink for whole layer

This table of interference gases primarily expresses the interference of each coexisting gas in the concentration range, that is equivalent to the gas concentration. Therefore, the test result may show positive results due to other substances not listed in the table. If more information is needed, please contact us or our distributors in your territory.

INSTRUCTIONS ON DISPOSAL :

The reagent of the primary tube uses a small amount of lead. The reagent of the analyser tube does not use toxic substances. When disposing the tube regardless of whether it has been used or not, follow the rules and regulations of your local government.

WARRANTY :

If you have any questions regarding gas detection and the quality of the tubes, please feel free to contact your Gastec representatives.

Manufacturer : Gastec Corporation

8-8-6 Fukayanaka, Ayase-City, Kanagawa 252-1195, Japan

<http://www.gastec.co.jp/>

Telephone +81-467-79-3910 Facsimile +81-467-79-3979

IM0076E2

Printed in Japan

12E1Z